



# राष्ट्रीय औषधीय शिक्षा तथा अनुसंधान संस्थान गुवाहाटी

## NATIONAL INSTITUTE OF PHARMACEUTICAL EDUCATION AND RESEARCH-GUWAHATI

(Department of Pharmaceuticals, Ministry of Chemicals and Fertilizers, Govt. of India)  
SilaKatamur (Halugurisuk), P.O.: Changsari, Dist: Kamrup, Assam, Pin: 781101.

No. NIPER-G/S&P/92/GAP-171/RE/2024-25/109

Date: 28/11/2024

### CORRIGENDUM

Subsequent to the pre-bid meeting held with prospective bidders on 25<sup>th</sup> November 2024 at 02:00 pm in reference to the GeM Bid no. **GEM/2024/B/5609867** for the supply and installation of **Rotary Evaporator (20 Litres) with Vacuum Pump and Circulating Chiller**, please note the following amendments:

Sl. No.	Original Tender Specifications as per the tender dated on 14/11/2024	Amended Tender Specification as per the pre-bid meeting dated on 25/11/2024
1.	<b>Pg. no. 03, Clause 5:</b> A detailed point wise technical compliance/deviation statement to be provided enabling user department to have clear idea of technical compliance of offer submitted.	<b>Pg. no. 03, Clause 5:</b> A detailed point wise technical compliance/deviation statement to be provided enabling user department to have clear idea of technical compliance of offer submitted. In case of No-deviations, indicate that the tender requirements are 100% fulfilled and no deviation to report.
2.	<b>Pg. no. 03, Clause 4.6:</b> The bidder should have supplied and installed similar kind of Items to Centrally Funded Technical Institutions / National Laboratories/Research Institutes/Medical Institutes (CSIR labs, IITs, NIPERs, AIIMS, NITs, IISCs, IISERs, IIITs etc.) and should provide Purchase Orders of those as documentary evidence.	<b>Pg. no. 03, Clause 4.6:</b> The bidder should have supplied and installed similar kind of Items to Centrally Funded Technical Institutions / National Laboratories/Research Institutes/Medical Institutes (CSIR labs, IITs, NIPERs, AIIMS, NITs, IISCs, IISERs, IIITs etc.) and should provide Purchase Orders of those as documentary evidence (the value of the order may be concealed).
3.	<b>Pg. no. 06, (Under Rotary Evaporator Unit)</b> Features and Safety	
4.	<b>Pg. no. 06, (Under Rotary Evaporator Unit)</b> Rotation Speed: Electronically controlled 6-160 rpm or better with digital display	<b>Pg. no. 06, (Under Rotary Evaporator Unit)</b> Rotation Speed: Electronically controlled 5-150 rpm or better with digital display
5.	<b>Pg. no. 06, (Under Rotary Evaporator Unit)</b> Control panel degree of protection: IP45 or better	<b>Pg. no. 06, (Under Rotary Evaporator Unit)</b> Control panel degree of protection: IP20 or better
6.	<b>Pg. no. 07, (Under Vacuum Pump)</b> Protection Class: IP40 or better	<b>Pg. no. 07, (Under Vacuum Pump)</b> Protection Class: IP20 or better

7.	<u>Addendum</u>	Rotary Evaporator, Vacuum Pump and Circulating Chiller must be of the same OEM.
		Unloading at NIPER-Guwahati premises and safe shifting of the item to identified location, unpacking etc. is suppliers' responsibility.

All other terms and conditions mentioned in the tender document remains unaltered.

This is for information of all concerned.



Stores & Purchase Officer  
NIPER-Guwahati